



Polish Centre for Testing and Certification

BM.433.0352.2021.1262.22.EP

Warsaw, 09.08.2022

To Whom it May Concern,

Polish Centre for Testing and Certification declares, that the information about reference numbers in the EC Certificate no. **1434-IVDD-225/2022** has been updated – added new brand Spring SARS-CoV-2 Antigen Saliva B Test with the reference numbers: SP-SL502-1B, SP-SL502-3B, SP-SL502-5B, SP-SL502-15B, SP-SL502-20B.

Implementation of the change does not represent a significant change in design or intended purpose under IVDR 2017/746 Article 110(3) and the related IVDD Certificate no. **1434-IVDD-225/2022** issued 25.05.2022 remains valid until 27.05.2025.

Yours Sincerely,

Elektronicznie podpisany
przez Tomasz Artur Koeber
Data: 2022.08.10 07:56:51
+02'00'

Director
Medical Devices Certification Department

Polish Centre for Testing and Certification
469 Puławska Street, 02-844 Warsaw
Tel.: +48 22 46 45 200, Fax: +48 22 46 45 251
pcbc@pcbc.gov.pl

Branch PCBC in Gdansk
18A Jakuba Wejhera Street, 80-346 Gdańsk
Tel.: +48 663 130 693
gdansk@pcbc.gov.pl

Branch PCBC in Pila
11 Sniadeckich Street, 64-920 Pila
Tel.: +48 67 21 38 200, Fax: +48 67 21 38 384
pila@pcbc.gov.pl

NIP: 951-20-63-356

REGON: 015276609

KRS: 0000144813

INITIAL CAPITAL: 16.000.000 PLN (FULLY PAID)